



Ref No. : 004383

Date: October 9, 2014

SUBJECT : Required Balance Specification for Gold, Diamond, Perfume and Consumable Product

Dear Valuable Shop Owner,

In compliance to **OIML International Standard** and as per **ESMA requirements**, GOLD, DIAMOND, PERFUME AND CONSUMABLE PRODUCT balances that are being used in the market shall fully comply as explained below. Table below that is derived from UAE.S/GSO OIML R76-1:2009 standard describes the minimum requirement for the balance classification / accuracy.

Table 3-1 of UAE.S/GSO OIML R76-1:2009

S/N	Classification of Balance	Capacity	Readability	Usage
1	I	Full Range	0.001g	Diamond
2	II	Full Range	0.01 g	Gold / Jewellery / Perfume
3	III	Up to 1 Kg	1 g	Nuts, Salt, Spices, Chocolates, other similar products
4		More than 1 Kg	2 g	
5		Up to 30 Kg		
6	III	Up to 30 Kg	5 g	Fish , Vegetables, Fruits, Sugar, Rice, Meat, other similar products
7		More than 30 Kg	10 g	
8		Up to 100 Kg		
9		More than 100 Kg	0.1 % from the Full Capacity	
10		Up to 20 Ton		

To further assist you, we have a list of recommended commercial balances stating the manufacturer, model number, instrument name, distributor in UAE and its classification to serve as a guide for you. (Refer to Attachment – English and Arabic)

In case your electronic balance is out of tolerance as per the requirements of ESMA & OIML standard, your balance will be treated as **UNFIT BALANCE** and you are legally required to remove from service or otherwise you can be legally penalized for violating the laws & regulations.

For further clarifications or any inquiries, please contact us on:

Tel No. : 04-3027171
Fax No. : 04-3027203
E-mail : CALIB@dm.gov.ae

Thanks and best regards.

Eng. Fareed Mohammad Ismail
Head of Metrology Section

Dubai Central Laboratory Department – Dubai Municipality